

**Organisation Specific (Patient Safety Alerts & Notices) Compliance position for Cwm Taf Morgannwg University Health Board
As at 16/02/2022**

There are 2 outstanding alerts, and 7 outstanding notices

PSA No	Date of Issue	NHS WALES - ALERTS	Compliance Deadline Date	Non-compliant /No response	All Wales Compliance Position	DU comments Any related NEs/SIs or Themes
PSA008	25/05/2017	Nasogastric tube misplacement: continuing risk of death and severe harm	30/11/2017	Non-compliant	5 of 8 compliant	Four years overdue. ALL-Wales pH STRIP ISSUE BEING ADDRESSED.
PSA014	12/10/2021	Inappropriate anticoagulation of patients with a mechanical heart valve	28/10/2021	No response	5 of 8 compliant	
PSN No	Date of Issue	NHS WALES - NOTICES	Compliance Deadline Date	Non-compliant /No response	All Wales Compliance Position	DU comments Any related NEs/SIs or Themes
PSN051	28/02/2020	Depleted batteries in intraosseous injectors	28/08/2020	Non-compliant	5 of 7 compliant	One year overdue.
PSN055	21/10/2020	THIS NOTICE REPLACES PSN030 Safe Storage of Medicines: Cupboards	30/09/2021	No response	6 of 10 compliant	
PSN056	27/10/2020	Foreign Body Aspiration during intubation, advanced airway management or ventilation	01/07/2021	No response	7 of 9 compliant	
PSN058	13/07/2021	Urgent assessment/treatment following ingestion of 'super strong' magnets	13/10/2021	No response	3 of 8 compliant	
PSN059	10/09/2021	Eliminating the risk of inadvertent connection to medical air via a flowmeter	16/12/2021	Non-compliant	5 of 7 compliant	
PSN060	27/09/2021	Reducing the risk of inadvertent administration of oral medication by the wrong route	20/12/2021	No response	5 of 9 compliant	

PSA/N No.	Compliance deadline date	TITLE	CHANGES TO NOTE SINCE 17/01/2022
PSN057	31/01/2022	Emergency Steroid Therapy Cards: Supporting Early Recognition & Management of Adrenal Crisis in Adults and Children	26/01/2022 NON-COMPLIANT.