



GIG
CYMRU
NHS
WALES

Bwrdd Iechyd Prifysgol
Cwm Taf Morgannwg
University Health Board



Freedom of Information Request: Our Reference CTMUHB_55_25

You asked:

Under the Freedom of Information Act, I would like to request the following information:

In the last quarter, what percentage of the following flexible endoscopes:

- Gastroscopes
- Colonoscopes
- Duodenoscopes

were returned to the manual cleaning process because they were not used within 3 hours after removal from the endoscope washer disinfectant (EWD) or automatic endoscope reprocessor, as per BSG and ESGE guidelines.

Our response:

Please see our response provided in the table below as captured.

HOSPITAL & SCOPE	Time Exceeded	Total Number processed
Prince Charles Duodenal	65	115
Prince Charles Colonoscopes	61	722
Prince Charles Gastroscopes	72	902
Royal Glamorgan Duodenal	17	85
Royal Glamorgan Colonoscopes	70	728
Royal Glamorgan Gastroscopes	72	587

Please note the data above does not include the Princess of Wales Hospital, as unfortunately due to disparate systems they are unable to easily extract this data.