



GIG  
CYMRU  
NHS  
WALES

Bwrdd Iechyd Prifysgol  
Cwm Taf Morgannwg  
University Health Board



## Freedom of Information Request: Our Reference CTMUHB\_278\_25

### You asked:

Under the Freedom of Information Act, I would like to request the following information:

**1. Is the endoscopy decontamination unit a single-room or double-room unit?**

Single-sided Endoscope Washer-Disinfectors (EWDs).

**a. If it is a single-room decontamination unit, what is the average and maximum temperature recorded, within the standard working hours?**

Not recorded, this is via plant room ventilation (staff do have temp controlled system).

**b. If it is a double-room decontamination unit, what is the average and maximum temperature recorded in both the clean and dirty room, within the standard working hours?**

N/A.

**2. Which process challenge device is currently being used?**

Independent contractors complete weekly testing using-GETINGE Pro CD – with Getinge equipment., must be compatible with EWD's QT-/Annual testing Getinge standards comply with 1MM/2MM/4MM Edinburgh test soil CD.

**3. Would the endoscopy unit be interested in trialling a new process challenge device?**

No.

**a. If yes, please provide a contact name and details.**

N/A.